

31

Notice of Allowability	Application No.	Applicant(s)	
	10/528,900	MATSUDA, TOMONOBU	
	Examiner	Art Unit	
	Sang Nguyen	2877	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address--

All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. **THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS.** This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.

1. ☒ This communication is responsive to 8/18/06.
2. ☒ The allowed claim(s) is/are 1-20.
3. ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
 - a) ☐ All b) ☐ Some* c) ☐ None of the:
 1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)).

* Certified copies not received: _____.

Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application.
THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.

4. ☐ A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.
5. ☐ CORRECTED DRAWINGS (as "replacement sheets") must be submitted.
 - (a) ☐ including changes required by the Notice of Draftsperson's Patent Drawing Review (PTO-948) attached
 - 1) ☐ hereto or 2) ☐ to Paper No./Mail Date _____.
 - (b) ☐ including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date _____.

Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).
6. ☐ DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.

Attachment(s)

- | | |
|---|---|
| 1. <input type="checkbox"/> Notice of References Cited (PTO-892) | 5. <input type="checkbox"/> Notice of Informal Patent Application (PTO-152) |
| 2. <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | 6. <input checked="" type="checkbox"/> Interview Summary (PTO-413),
Paper No./Mail Date <u>8/18/06</u> . |
| 3. <input type="checkbox"/> Information Disclosure Statements (PTO-1449 or PTO/SB/08),
Paper No./Mail Date _____ | 7. <input checked="" type="checkbox"/> Examiner's Amendment/Comment |
| 4. <input type="checkbox"/> Examiner's Comment Regarding Requirement for Deposit
of Biological Material | 8. <input checked="" type="checkbox"/> Examiner's Statement of Reasons for Allowance |
| | 9. <input type="checkbox"/> Other _____. |

EXAMINER'S AMENDMENT

An examiner's amendment to the record appears below. Should the changes and/or additions be unacceptable to applicant, an amendment may be filed as provided by 37 CFR 1.312. To ensure consideration of such an amendment, it **MUST** be submitted no later than the payment of the issue fee.

Authorization for this examiner's amendment was given in a telephone interview with Mr. Joseph P. Carrier (Reg. No. 31,748) on 08/18/06.

The application has been amended as follows:

16. (New) A particle measuring apparatus comprising:

a flow cell having a first passage and second passage extending continuously from said first passage;

a particle monitoring area formed in the first passage by irradiating the flow cell with light; and

a plurality of condensers which condense light scattered by particles contained in sample fluid passing through the particle monitoring area so as to obtain information including diameter of the particles,

wherein a central axis of the passage substantially corresponds to an optical axis of the condensers and inner walls of the flow cell are arranged so as not to impede the scattered light from entering an outmost peripheral portions of the condensers; **and**

said condensers are provided on opposite sides of the flow cell; and

a width of the second passage is greater than a width of the first passage so as not to impede the scattered light from entering an outmost peripheral portions of the condensers.

The following is an examiner's statement of reasons for allowance:

As to independent claims 1, 12, and 16 are allowable over the prior art for at least the reason that the prior art of record, taken alone or in combination, fails discloses or render obvious a particle measuring apparatus comprising all the specific elements with the specific combination including of a width of the second passage is greater than a width of the first passage so as not to impede the scattered light from entering an outmost peripheral portions of the condensers, wherein a flow cell having a first passage and second passage extending continuously from said first passage in combination with the rest of the limitation of claims 1, 12, and 16. The dependent claims 2-11, 13-15, and 17-20 are allowed by virtue of their dependence on claims 1, 12, and 16.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."


Any inquiry concerning this communication or earlier communications from the examiner should be directed to Sang Nguyen whose telephone number is (571) 272-2425. The examiner can normally be reached on 9:30 am to 7:00 pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Gregory J. Toatley, Jr. can be reached on (571) 272-2800 ext. 77. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

August 18, 2006

Sang Nguyen 
Patent Examiner


Gregory J. Toatley, Jr.
Supervisory Patent Examiner
Art Unit 2877
Technology Center 2800
18 AUG 06